

### RAPID TESTING TEMPERATURE LOG

Thermometer location

Month/Year

Acceptable temperature range\*

Day	Initials	High Temp	Low Temp	Corrective action taken when temperature is out of range
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				
31				

\*The acceptable range for **Determine test kit** storage is 2 to 30° C or 36 to 86° F and the acceptable range for control storage is 2 to 8° C or 36 to 46° F. The acceptable range for the **Rapid Syphilis Health Check Test** is 4 to 30° C or 39 to 86° F and the acceptable range for syphilis control storage is 2 to 8°C or 36 to 46° F.

Reviewed by (Print Name)

Date Reviewed